


# Pipette Tips and Filter Tips in Standard and Ultra Low Retention Quality




Optimum results are achieved in combination with BRAND microliter pipettes. BRAND quality tips seat perfectly for precise analyses.

Pipette and tips from BRAND are tested for BRAND pipettes and most of the pipette types of GILSON®, Thermo Fisher Scientific FINNPIPETTE®, Eppendorf® and BIOHIT®/sartorius®.

The 5 ml tip is exclusively tested for BRAND pipettes and Thermo Fisher Scientific FINNPIPETTE®. The 10 ml tip is tested for BRAND and Eppendorf®.



Note: Pipette shafts can be subject to modification. The fit depends on the manufacturer, pipette type, serial number, and date of manufacture, among other things. We recommend checking the fit of the tips using the free sample bag.



## The right Pipette Tip

Pipette tips Filter tips Volume range	Transferpette® single channel Nominal volume														Transferpette® multichannel Nominal volume											
	1 µl	2.5 µl	5 µl	10 µl	10 µl	20 µl	20 µl	25 µl	50 µl	100 µl	200 µl	200 µl	250 µl	500 µl	1000 µl	2 ml	5 ml	10 ml	10 µl	20 µl	25 µl	50 µl	100 µl	200 µl	300 µl	
0.1 - 20 µl	✓	✓ <sup>2)</sup>		✓ <sup>2)</sup>															✓	✓						
0.5 - 20 µl		✓		✓		✓													✓	✓						
1 - 50 µl		✓ <sup>2)</sup>		✓		✓													✓	✓						
2 - 200 µl <sup>1)</sup>			✓		✓		✓	✓	✓	✓	✓										✓	✓	✓	✓	✓	✓
5 - 300 µl			✓		✓		✓	✓	✓	✓	✓										✓	✓	✓	✓	✓	✓
50 - 1000 µl <sup>1)</sup>												✓	✓	✓	✓											
0.5 - 5 ml																✓	✓									
1 - 10 ml																		✓								
0.1 - 1 µl	✓	✓ <sup>2)</sup>		✓ <sup>2)</sup>															✓							
0.5 - 10 µl		✓		✓		✓													✓	✓						
1 - 20 µl				✓		✓													✓	✓						
2 - 20 µl			✓		✓		✓	✓	✓	✓	✓										✓	✓	✓	✓	✓	✓
5 - 100 µl			✓		✓		✓	✓	✓	✓	✓										✓	✓	✓	✓	✓	✓
5 - 200 µl											✓										✓	✓	✓	✓	✓	✓
50 - 1000 µl														✓	✓											

✓ = Tip volume less than pipette's nominal volume

<sup>1)</sup> Bulk tips are yellow or blue colored, racked tips are colorless in a yellow-colored or blue-colored tip tray

<sup>2)</sup> Calibration and possible adjustment needed