# MDBK Serum-reduced Medium (90-080)

### Instructions for Use

## CORNING

#### **Media Basics**

Dry powder MDBK Serum-reduced Medium is phenol red-free powder medium, without animal-derived components, with L-glutamine, without sodium bicarbonate. It is mainly used for MDBK cells growth with serum-reduced suspension culture and virus production.

#### **Specifications**

Description	Cat. No.	Quantity	<b>Expiration Date</b>
MDBK Serum-reduced Medium	90-080-PBT	1 x 10L	2 years
	90-080-PCT	1 x 50L	2 years
	90-080-PDT	1 x 100L	2 years

1L of media is produced using 17.65 g of dry powder medium.

#### Storage Conditions

2°C to 8°C, sealed and protected from light.

#### **Preparation Method (1L)**

- 1. Check the condition of the powder before preparation to ensure it flows freely and is free of clumps. Clumping indicates the medium has absorbed moisture, which may impact product quality and stability.
- 2. Add room temperature (20°C to 30°C) cell culture grade water (Corning 25-055-CM) or ultrapure water to a clean container to 90% of the final volume (~900 mL). Place water on a stir plate and rotate at slow setting.
- 3. Carefully weigh out 17.65 g of dry powder medium onto a clean weighing dish. Slowly add the medium powder to the container, rinse the weighing dish with water and add to container. Stir the solution for 30 minutes until the bulk of the dry powder medium has dissolved into solution. **NOTE:** Do not add medium powder first and then water, as it will clump and prevent full dissolution.

**NOTE:** Unused portion of the powdered media should be placed in an air tight container and stored at conditions listed above.

- 4. Adjust the pH value to 8.50-8.70 by slowly adding 5N NaOH or 5N HCl and stir solution for an additional 30 min. to ensure all components are fully dissolved.
- 5. Add 2.0 g of sodium bicarbonate (Corning 61-065-RO).
- 6. Stir continuously for 10 minutes until completely dissolved, then add water to within 1% of the final volume.
- 7. Adjust the pH value to 7.0-7.5 by slowly adding either 5N NaOH solution or 5N HCl and stir until the desired pH is stable.
- 8. Stop stirring, add water to 1L, and filter sterilize using a 0.22 μm or smaller filter (Corning 431098).
- 9. For final media preparation, add Fetal Bovine Serum (FBS) to desired final concentration and filter sterilized if desired.
- 10. The prepared medium should be stored at 2-8°C and protected from light.

NOTE: There are 2 methods to prepare 10L, 50L, and 100L quantities:

- 1. 17.65 g of powder can be weighed per liter and be prepared according to the method above.
- 2. A single package (10L, 50L, and 100L) can be completely prepared at one time according to the method above. 2.0 g/L of sodium bicarbonate should be added in proportion to the volume of media prepared.

Warranty/Disclaimer: Unless otherwise specified, all products are for research use or general laboratory use only.\* Not intended for use in diagnostic or therapeutic procedures. Not for use in humans. These products are not intended to mitigate the presence of microorganisms on surfaces or in the environment, where such organisms can be deleterious to humans or the environment. Corning Life Sciences makes no claims regarding the performance of these products for clinical or diagnostic applications. \*For a listing of US medical devices, regulatory classifications or specific information on claims, visit www.corning.com/resources.

Corning's products are not specifically designed and tested for diagnostic testing. Many Corning products, though not specific for diagnostic testing, can be used in the workflow and preparation of the test at the customers discretion. Customers may use these products to support their claims. We cannot make any claims or statements that our products are approved for diagnostic testing either directly or indirectly. The customer is responsible for any testing, validation, and/or regulatory submissions that may be required to support the safety and efficacy of their intended application.

## CORNING

Corning Incorporated Life Sciences

www.corning.com/lifesciences

NORTH AMERICA t 800.492.1110

t 978.442.2200 ASIA/PACIFIC

#### Australia/New Zealand t 61 427286832 Chinese Mainland

t 86 21 3338 4338

India t 91 124 4604000 Japan t 81 3-3586 1996 Korea t 82 2-796-9500 Singapore t 65 6572-9740 Taiwan t 886 2-2716-0338 EUROPE CSEurope@corning.com France t 0800 916 882 Germany t 0800 101 1153 The Netherlands t 020 655 79 28 United Kingdom t 0800 376 8660 **All Other European Countries** t +31 (0) 206 59 60 51

#### LATIN AMERICA

grupoLA@corning.com Brazil t 55 (11) 3089-7400 Mexico t (52-81) 8158-8400