

Specification Sheet

Specification Number: 8920H-450S Revision: 02 Date Issued: 25-Mar-20

1. Part Number: 8920H-450S

2. Photograph:





3. Product Information:

Description: SteriWare LiquiDip Mono 450mm - Sterile

Use: Single use, disposable sampler

Material of Construction: High Density Polyethylene (HDPE) - virgin

FDA acceptable CFR 177.1520EU Regulations 10/2011 compliant

• EC regulation 1935/2004 compliant

• BSE/TSE free

Method of Construction: Ultrasonic welding

Overall Length 430mm

Maximum Diameter: 26mm

Outer Diameter of Tube: 21mm

Diameter of hole in tip: 6mm

Nominal Weight of Sampler: 50g

Moulding Environment: Class 100,000 Cleanroom

Individually Bagged? Yes (heat sealed PE/PET bag)
Method of Sterilisation: Gamma Irradiation (25-45 kGy)

Quantity per Box: 20

Recommended Storage Conditions: Dry and ambient temperature

Shelf Life: 5 years from the date of manufacture

ATEX: This product has not been ATEX tested

4. Change Control: Sampling Systems operates a policy of continual improvement.

To be kept informed of all major changes to any product in the SteriWare range please email your contact details to

change.control@sampling.com



4-5 Forge Mills Park, Coleshill. Warwickshire. B46 1JH. UK. Tel: +44(0)1675 466992 Email: info@sampling.com Web: www.sampling.com



Specification Sheet

Specification Number:8920H-450SRevision:02Date Issued:25-Mar-20

5. Country of Origin: Made and packed in the UK.

6. Repeated Use: This is a single-use, disposable device. Repeated use will lead to

failure.

7. Approval: Signed:

Name: Kuldeep Flora

Quality Manager

Date: 25-Mar-20

Position: