

## Specification Sheet

Specification Number: 8064H-500 Revision: 08 Date Issued: 29-Jun-21

**1. Part Number:** 8064H-500

2. Photograph:

## 3. Product Information:

Description: SteriWare® Powder Probe 500mm

Use: Single use powder sampler

Material of Construction: High Density Polyethylene (HDPE) – virgin

Conforms to FDA 21 CFR 177.1520
Conforms to EU Regulations 10/2011
Conforms to EC Regulations 1935/2004

• BSE/TSE free

Method of Construction: Ultrasonic welding

Nominal Length: 615mm
Blade Length: 497mm
Diameter of Blade: 22mm
Width of opening in blade: 10mm
Nominal Weight of Sampler: 89g

Moulding & Primary Packing Environment: Class 100,000 Cleanroom Individually Bagged? Yes (heat sealed PE/PET bag)

Method of Sterilisation: N/A
Quantity per Box: 100

Recommended Storage Conditions: Dry and ambient temperature

Shelf Life: 5 years from the date of manufacture

**4. Change Control:** Sampling Systems operates a policy of continual improvement.

To be kept informed of all major changes to any product in the

SteriWare range please email your contact details to

change.control@sampling.com

**5. Country of Origin:**Made and packed in the UK.



4-5 Forge Mills Park, Coleshill. Warwickshire. B46 1JH. UK. Tel: +44(0)1675 466992 Email: info@sampling.com Web: www.sampling.com



## **Specification Sheet**

Specification Number: 8064H-500 Revision: 08 Date Issued: 29-Jun-21

**6. Repeated Use:**This is a single-use, disposable sampler. Repeated use will lead to failure.

**7. Approval:** Signed:

Name: Kuldeep Flora
Position: Quality Manager

Date: 29-Jun-21

